



CERTIFICATE OF COMPLIANCE (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Braun GmbH Industrie-Elektronik
Address	Esslinger Straße 26, DE 71334 Waiblingen Germany
Project Number	LR 1323-1R1
Product	Sensor
Model Number	Type A5S0***** ** (OrdLoc) Type A5S1***** ** (HazLoc) Type A5S1*****-n** (HazLoc) Note: The A5S Series sensor has a number of options defined by the full Type coding (see report)
Ratings	Canada: Sensor type A5S1***** ** Class I, Div 1, Groups A, B, C, D, Temperature code T6/T4, 'Intrinsically safe' Class I, Zone 0, Ex ia IIC T6/T4 Ga Sensor type A5S1*****-n** Class I, Div 2, Groups A, B, C, D, Temperature code T6/T4, 'Non-incendive' Class I, Zone 2, Ex nA IIC T6/T4 Gc US: Sensor type A5S1***** ** Class I, Div 1, Groups A, B, C, D, Temperature code T6/T4, 'Intrinsically safe' Class I, Zone 0, AEx ia IIC T6/T4 Sensor type A5S1*****-n** Class I, Div 2, Groups A, B, C, D, Temperature code T6/T4, 'Non-incendive' Class I, Zone 2, AEx nA IIC T6/T4 Note: complete electrical ratings ("See instruction manual" – detailed in the instructions and can be identified from model number)
Applicable Standards	Canadian standards Canadian Electrical Code (CEC) dated 2012 CSA C22.2 No. 60079-0:11 CSA C22.2 No. 60079-11:11 CSA C22.2 No. 60079-15:12 CSA C22.2 No. 157-92, (R2012) CSA C22.2 No. 213-1987 (R2013) CSA-C22.2 No.61010-1-12, edition 3



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.
81 Kelfield St., Unit 8, Toronto, ON, M9W 5A3, Canada Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca





QPS Evaluation Services Inc.
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
LR 1323

US standards

National Electrical Code (NEC) dated 2014
 UL 60079-0 (12.00.01) -2013
 UL 60079-11 (12.02.01) -2013
 UL 60079-15 (12.12.02) -2012
 UL 913 – 8th edition
 ISA 12.12.01 – 2013
 UL-61010-1 (82.02,.01), edition 3

Factory/Manufacturing Location	Braun GmbH Industrie-Elektronik Esslinger Straße 26, DE 71334 Waiblingen Germany
--------------------------------	--

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: **D. Adams, P.Eng.**
 Certification Manager, Hazardous Locations Dept. [Ex Equipment]

Signature: 

Date: March 19, 2024



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.
 81 Kelfield St., Unit 8, Toronto, ON, M9W 5A3, Canada Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca

